



Head Office: 106 - 116 Epsom Road,
Rosebery NSW 2018 AUSTRALIA

Phone: 1300 780 078
Fax: (02) 9313 6444

International Phone: 1300 780 078
International Fax: (02) 9313 6444

Email: sales@livingstone.com.au
Website: www.livingstone.com.au

MATERIAL SAFETY DATA SHEET

PRODUCT CODE: NEBLIV

PAGE 1 OF 2

VALID TILL: 10/05/2016

Item No	Description
NEBLIV	ECON-O-MIST NEBULISER THERAPY PUMP BD EACH
Product:	ECON-O-MIST NEBULISER THERAPY PUMP BD EACH
Brand	AllerSearch

Thermal Cut Out Device

240V 50Hz 90W

Dimensions: (cm) 13.5H x 20D x 23W

Contents: 1 only Econ O Mist Aerosol Therapy Pump

Accessories: 1 only jet nebuliser, 1 only adult face mask, 1 only child face mask 1 only air tube
Strong, lightweight, impact resistant plastic case with carry handle
High Quality filtered air
Convenient storage compartment for mask and jet nebuliser

MATERIAL SAFETY DATA SHEET

PRODUCT CODE: NEBLIV

PAGE 2 OF 2

VALID TILL: 10/05/2016

Operating Instructions

- Connect power cable to 240V power point.
- Add medication into jet nebuliser and attach face mask to the outlet of jet nebuliser.
- Attach air tube to bottom of jet nebuliser and place mask on face.
- Switch on Econ-O-Mist and breath normally through the mouth.
- Switch off pump when nebulisation is completed or in between treatments.
- Maximum operating time is 30 minutes then allow 30 minutes for cooling.

Cleaning Instructions

- Dismantle the jet nebuliser and rinse in warm water after each use.
- Disinfect face mask and jet nebuliser components twice a week in a non phenolic disinfectant available from your pharmacy.

TROUBLE SHOOTING

Pump not operating

- Check power point with an appliance that you know is operating correctly.
- Check fuse.
- Allow 30 minutes cooling.
- If pump becomes noisy remove inlet filter and replace with new filter as soon as possible.
- Air inlet filter should be changed every 12 months, when nebulisation time takes longer or when the filter becomes discoloured.

Jet Nebuliser (if not misting or takes longer than normal)

- Check that the centre jet is not blocked.
- Twist and push down the central jet inside the nebuliser.
- Check inlet filter for blockage.
- If problem still exists please contact the supplier.

TECHNICAL FEATURES

Pump

- No lubrication and minimal maintenance required.
- ECON-O-MIST is manufactured to comply with Australian Standards AS 3200 (amend. 1990).
- 240V 50Hz 90W
- Dynamic Air Flow 6 litres per minute

Jet Nebuliser

- | | |
|------------------------|--|
| ▪ Recommended Air Flow | 5-10 litres per minute |
| ▪ Maximum Fill Volume | 10mL |
| ▪ Minimum Fill Volume | 2mL |
| ▪ Particle Size (MMD) | 3.9 microns @ 6 litres per minute air flow |

WARRANTY

Your ECON-O-MIST Aerosol Therapy Pump has a warranty of 24 months from date of original purchase. Warranty covers faulty workmanship and/or materials but does not cover normal wear and tear, incorrect use or accessories.

Warranty - 2 Years